


Report: COA Evaluation Summary

OLCC License No. 10087092BDA | ORELAP ID. 4147

545 SW 2nd Street, Corvallis OR. 97333 | 541.257.5002 | services@preelab.com | Preelab.com

For OLCC/OHA Compliance Purposes.

Product Description		Evaluation Summary		
Client:	Anthem Hemp	Moisture Analysis	Test Not Required	
Product Name:	CBG Isolate TST48			
Production Date:	2021/10/26	Cannabinoid Potency Analysis		
Expiry Date:	2023/10/26	Total THC *	Total CBD *	Abrv. Dry Wt. % Dry Wt. mg/g
Matrix:	Hemp Concentrate	0.00 %	0.00 %	THCA < LOQ < LOQ
Metrc Source ID:	n/a	0.0 mg/g	0.0 mg/g	Δ-9-THC < LOQ < LOQ
Metrc Package ID:	n/a	THCV < LOQ < LOQ		
License Number:	n/a	CBDA < LOQ < LOQ		
Date Collected:	2021-10-27	CBD < LOQ < LOQ		
Date Received:	2021-10-27	CBGA < LOQ < LOQ		
Report Date:	2021-10-29	CBG 100.00 % 1000.0 mg/g		
Report ID:	A4994-03	CBDVA < LOQ < LOQ		
Tests Requested:	Cannabinoid Potency Analysis Pesticide Analysis Residual Solvent Analysis	CBDV < LOQ < LOQ		
		CBN < LOQ < LOQ		
		CBL < LOQ < LOQ		
		CBC < LOQ < LOQ		
				
		Pesticide Analysis	Pesticide Status	
		Pass	No Pesticides Were Detected above Oregon's action limit as stated in OAR 333-007-0400.	

* moisture compensated & adjusted for the loss of carboxylic acid group - OAR 333-064-0100

Report: Case Narrative

This certificate of analysis is prepared for...

Anthem Hemp

18253 Gothard St. Huntington Beach, CA 92648

This report presents the analytical findings for the sample collected on 2021-10-27 by Skyler Smith using sampling plan A4994 and received by PREE Laboratory on 2021-10-27. The sample was assigned a laboratory ID of A4994-03. The results in this report only apply to sample A4994-03.

This report shall not be reproduced, except in full, without written consent of PREE Laboratory. Report alterations by any other entity beside PREE Laboratory are not allowed. If alterations are made to the original report after the initial release, they will be noted on the case narrative below.

The testing methods used are of sufficient sensitivity to meet the compliance criteria set in OAR 333-007. However, it is the responsibility of the client to utilize the data to comply with standards set in OAR 333-007.

All analyses were performed in accordance with PREE Laboratory's NELAP/TNI approved quality control system and all quality control data was within the laboratory's predefined acceptance criteria unless otherwise noted in the case narrative of this report. General comments are also recorded below.

Notes:

The Oregon Department of Agriculture requires hemp products to not contain more than 0.35% total THC, per OAR 603-048. Residual solvent analysis was subcontracted. The report from the subcontracting laboratory is attached. No special conditions were noted during the processing and testing of the sample.



Sardar, Tamzid M. | Laboratory Director
Corvallis, Oregon



If you have any questions regarding the information in this report, please feel free to call 541-257-5002 or email PREE at services@preelab.com.

Report: Evaluation Detail



OLCC License No. 10087092BDA | ORELAP ID. 4147

545 SW 2nd Street, Corvallis OR. 97333 | 541.257.5002 | services@preelab.com | Preelab.com

For OLCC/OHA Compliance Purposes.

Moisture Analysis	Evaluation Detail					
	Moisture Analysis	Test Not Requested/Required				
Cannabinoid Potency Analysis	Evaluation Detail					
Product Name: CBG Isolate TST48	Cannabinoid Potency Analysis	Compound	Abrv.	Dry Wt. (%)	Dry Wt. (mg/g)	RL (%)
Analysis Date: 2021-10-28	Total THC *	Tetrahydro-cannabinolic acid	THCA	< LOQ	< LOQ	0.5 %
Testing Batch ID: V1892,1887,1882,1876	0.00 %	Delta9 Tetrahydro-cannabinol	Δ-9-THC	< LOQ	< LOQ	0.5 %
Testing Method: LSOP #303 Cannabinoid Quantification	0.0 mg/g	Delta8 Tetrahydro-cannabinol	Δ-8-THC	< LOQ	< LOQ	0.5 %
		Tetrahydrocannabivarin	THCV	< LOQ	< LOQ	0.5 %
	Total CBD *	Cannabidiolic acid	CBDA	< LOQ	< LOQ	0.5 %
	0.00 %	Cannabidiol	CBD	< LOQ	< LOQ	0.5 %
	0.0 mg/g	Cannabigerolic acid	CBGA	< LOQ	< LOQ	0.5 %
		Cannabigerol	CBG	100.00 %	1000.0	0.5 %
		Cannabidivarinic acid	CBDVA	< LOQ	< LOQ	0.5 %
		Cannabidivarin	CBDV	< LOQ	< LOQ	0.5 %
		Cannabinol	CBN	< LOQ	< LOQ	0.5 %
		Cannabicyclol	CBL	< LOQ	< LOQ	0.5 %
		Cannabichromene	CBC	< LOQ	< LOQ	0.5 %

Note: Accreditation for Δ-8-THC, THCV, CBGA,CBG, CBDVA, CBDV, CBL, CBC, CBN is not offered by ORELAP and therefore are not accredited tests.

* moisture compensated & adjusted for the loss of carboxylic acid group - OAR 333-064-0100

Report: Evaluation Detail



OLCC License No. 10087092BDA | ORELAP ID. 4147

545 SW 2nd Street, Corvallis OR. 97333 | 541.257.5002 | services@preelab.com | Preelab.com

For OLCC/OHA Compliance Purposes.

Pesticide Analysis

Product Name: **CBG Isolate TST48 Prim**

Analysis Date: 2021-10-28

Testing Batch ID: V1888,1887

Testing Method: LSOP #307 Pesticides by LCMS/MS

Evaluation Detail

Pesticide Name	Tested Value (ppm)	Pass Criteria (ppm)	LOQ (ppm)	Status Pass/Unsatisfactory
Abamectin	< LOQ	0.50	0.20	Pass
Acephate	< LOQ	0.40	0.04	Pass
Acequinocyl	< LOQ	2.00	0.20	Pass
Acetamiprid	< LOQ	0.20	0.04	Pass
Aldicarb	< LOQ	0.40	0.04	Pass
Azoxystrobin	< LOQ	0.20	0.04	Pass
Bifenazate	< LOQ	0.20	0.04	Pass
Bifenthrin	< LOQ	0.20	0.20	Pass
Boscalid	< LOQ	0.40	0.04	Pass
Carbaryl	< LOQ	0.20	0.04	Pass
Carbofuran	< LOQ	0.20	0.04	Pass
Chlorantraniliprole	< LOQ	0.20	0.04	Pass
Chlorfenapyr	< LOQ	1.00	0.20	Pass
Chlorpyrifos	< LOQ	0.20	0.04	Pass
Clofentezine	< LOQ	0.20	0.20	Pass
Cyfluthrin	< LOQ	1.00	1.00	Pass
Cypermethrin	< LOQ	1.00	1.00	Pass
Daminozide	< LOQ	1.00	0.20	Pass
Diazinon	< LOQ	0.20	0.04	Pass
Dichlorvos	< LOQ	1.00	0.20	Pass
Dimethoate	< LOQ	0.20	0.04	Pass
Ethoprophos	< LOQ	0.20	0.04	Pass
Etofenprox	< LOQ	0.40	0.20	Pass
Etoxazole	< LOQ	0.20	0.04	Pass
Fenoxycarb	< LOQ	0.20	0.04	Pass
Fenpyroximate	< LOQ	0.40	0.20	Pass
Fipronil	< LOQ	0.40	0.04	Pass
Flonicamid	< LOQ	1.00	0.04	Pass
Fludioxonil	< LOQ	0.40	0.20	Pass
Hexythiazox	< LOQ	1.00	0.04	Pass
Imazalil	< LOQ	0.20	0.04	Pass
Imidacloprid	< LOQ	0.40	0.04	Pass
Kresoxim-methyl	< LOQ	0.40	0.20	Pass

Continued on next page...

Report: Evaluation Detail



OLCC License No. 10087092BDA | ORELAP ID. 4147

545 SW 2nd Street, Corvallis OR. 97333 | 541.257.5002 | services@preelab.com | Preelab.com

For OLCC/OHA Compliance Purposes.

Pesticide Analysis

Evaluation Detail

Pesticide Name	Tested Value (ppm)	Pass Criteria (ppm)	LOQ (ppm)	Status Pass/Unsatisfactory
Malathion	< LOQ	0.20	0.04	Pass
Metalaxyl	< LOQ	0.20	0.04	Pass
Methiocarb	< LOQ	0.20	0.04	Pass
Methomyl	< LOQ	0.40	0.04	Pass
Methyl-Parathion	< LOQ	0.20	0.20	Pass
MGK-264	< LOQ	0.20	0.20	Pass
Myclobutanil	< LOQ	0.20	0.20	Pass
Naled	< LOQ	0.50	0.04	Pass
Oxamyl	< LOQ	1.00	0.04	Pass
Paclobutrazol	< LOQ	0.40	0.04	Pass
Permethrins	< LOQ	0.20	0.20	Pass
Phosmet	< LOQ	0.20	0.04	Pass
Piperonyl butoxide	< LOQ	2.00	0.04	Pass
Prallethrin	< LOQ	0.20	0.20	Pass
Propiconazole	< LOQ	0.40	0.20	Pass
Propoxur	< LOQ	0.20	0.04	Pass
Pyrethrins	< LOQ	1.00	1.00	Pass
Pyridaben	< LOQ	0.20	0.04	Pass
Spinosad	< LOQ	0.20	0.20	Pass
Spiromesifen	< LOQ	0.20	1.00	Pass
Spirotetramat	< LOQ	0.20	0.04	Pass
Spiroxamine	< LOQ	0.40	0.04	Pass
Tebuconazole	< LOQ	0.40	0.04	Pass
Thiacloprid	< LOQ	0.20	0.04	Pass
Thiamethoxam	< LOQ	0.20	0.04	Pass
Trifloxystrobin	< LOQ	0.20	0.04	Pass

Report: Quality Check



OLCC License No. 10087092BDA | ORELAP ID. 4147

545 SW 2nd Street, Corvallis OR. 97333 | 541.257.5002 | services@preelab.com | Preelab.com

For OLCC/OHA Compliance Purposes.

Moisture Analysis	Quality Control Detail						
	Moisture Analysis						
Cannabinoid Potency Analysis	Quality Control Detail						
Analysis Date: 2021-10-28	Cannabinoid Potency Analysis		MB	LCS	Expected Value (%)	Tested Value (%)	Pass Criteria
Testing Batch ID: V1892,1887,1882,1876	Tetrahydro-cannabinolic acid		○		< 0.1%	Not Used	< 0.1%
	Delta9 Tetrahydro-cannabinol		○		< 0.1%	Not Used	< 0.1%
	Cannabidiolic acid		○		< 0.1%	Not Used	< 0.1%
	Cannabidiol		○		< 0.1%	Not Used	< 0.1%
	Tetrahydro-cannabinolic acid			●	100.0%	Not Used	80-120%
	Delta9 Tetrahydro-cannabinol			●	100.0%	Not Used	80-120%
	Cannabidiolic acid			●	100.0%	Not Used	80-120%
	Cannabidiol			●	100.0%	Not Used	80-120%

Report: Quality Check



OLCC License No. 10087092BDA | ORELAP ID. 4147

545 SW 2nd Street, Corvallis OR. 97333 | 541.257.5002 | services@preelab.com | Preelab.com

For OLCC/OHA Compliance Purposes.

Pesticide Analysis

Analysis Date: 2021-10-28

Testing Batch ID: V1888,1887

Quality Control Detail

Pesticide Name	MB	Expected Value (ppm)	Tested Value (ppm)	Pass Criteria (ppm)
Abamectin	o	< 0.1	< 0.1	< 0.1
Acephate	o	< 0.02	< 0.02	< 0.02
Acequinocyl	o	< 0.1	< 0.1	< 0.1
Acetamiprid	o	< 0.02	< 0.02	< 0.02
Aldicarb	o	< 0.02	< 0.02	< 0.02
Azoxystrobin	o	< 0.02	< 0.02	< 0.02
Bifenazate	o	< 0.02	< 0.02	< 0.02
Bifenthrin	o	< 0.1	< 0.1	< 0.1
Boscalid	o	< 0.02	< 0.02	< 0.02
Carbaryl	o	< 0.02	< 0.02	< 0.02
Carbofuran	o	< 0.02	< 0.02	< 0.02
Chlorantraniliprole	o	< 0.02	< 0.02	< 0.02
Chlorfenapyr	o	< 0.1	< 0.1	< 0.1
Chlorpyrifos	o	< 0.02	< 0.02	< 0.02
Clofentezine	o	< 0.1	< 0.1	< 0.1
Cyfluthrin	o	< 0.5	< 0.5	< 0.5
Cypermethrin	o	< 0.5	< 0.5	< 0.5
Daminozide	o	< 0.1	< 0.1	< 0.1
Diazinon	o	< 0.02	< 0.02	< 0.02
Dichlorvos	o	< 0.1	< 0.1	< 0.1
Dimethoate	o	< 0.02	< 0.02	< 0.02
Ethoprophos	o	< 0.02	< 0.02	< 0.02
Etofenprox	o	< 0.1	< 0.1	< 0.1
Etoxazole	o	< 0.02	< 0.02	< 0.02
Fenoxycarb	o	< 0.02	< 0.02	< 0.02
Fenpyroximate	o	< 0.1	< 0.1	< 0.1
Fipronil	o	< 0.02	< 0.02	< 0.02
Flonicamid	o	< 0.02	< 0.02	< 0.02
Fludioxonil	o	< 0.1	< 0.1	< 0.1
Hexythiazox	o	< 0.02	< 0.02	< 0.02
Imazalil	o	< 0.02	< 0.02	< 0.02
Imidacloprid	o	< 0.02	< 0.02	< 0.02
Kresoxim-methyl	o	< 0.1	< 0.1	< 0.1

Continued on next page...

Report: Quality Check



OLCC License No. 10087092BDA | ORELAP ID. 4147

545 SW 2nd Street, Corvallis OR. 97333 | 541.257.5002 | services@preelab.com | Preelab.com

For OLCC/OHA Compliance Purposes.

Pesticide Analysis

Quality Control Detail

Pesticide Name	MB	Expected Value (ppm)	Tested Value (ppm)	Pass Criteria (ppm)
Malathion	o	< 0.02	< 0.02	< 0.02
Metalaxyl	o	< 0.02	< 0.02	< 0.02
Methiocarb	o	< 0.02	< 0.02	< 0.02
Methomyl	o	< 0.02	< 0.02	< 0.02
Methyl-Parathion	o	< 0.1	< 0.1	< 0.1
MGK-264 I	o	< 0.1	< 0.1	< 0.1
MGK-264 II	o	< 0.1	< 0.1	< 0.1
Myclobutanil	o	< 0.1	< 0.1	< 0.1
Naled	o	< 0.02	< 0.02	< 0.02
Oxamyl	o	< 0.02	< 0.02	< 0.02
Paclobutrazol	o	< 0.02	< 0.02	< 0.02
Permethrin - trans	o	< 0.1	< 0.1	< 0.1
Permethrin - cis	o	< 0.1	< 0.1	< 0.1
Phosmet	o	< 0.02	< 0.02	< 0.02
Piperonyl butoxide	o	< 0.02	< 0.02	< 0.02
Prallethrin	o	< 0.1	< 0.1	< 0.1
Propiconazole	o	< 0.1	< 0.1	< 0.1
Propoxur	o	< 0.02	< 0.02	< 0.02
Pyrethrin - Cinerin	o	< 0.5	< 0.5	< 0.5
Pyrethrin - Jasmolin	o	< 0.5	< 0.5	< 0.5
Pyrethrin - Pyrethrins	o	< 0.5	< 0.5	< 0.5
Pyridaben	o	< 0.02	< 0.02	< 0.02
Spinosyn A	o	< 0.1	< 0.1	< 0.1
Spinosyn D	o	< 0.1	< 0.1	< 0.1
Spiromesifen	o	< 0.5	< 0.5	< 0.5
Spirotetramat	o	< 0.02	< 0.02	< 0.02
Spiroxamine	o	< 0.02	< 0.02	< 0.02
Tebuconazole	o	< 0.02	< 0.02	< 0.02
Thiacloprid	o	< 0.02	< 0.02	< 0.02
Thiamethoxam	o	< 0.02	< 0.02	< 0.02
Trifloxystrobin	o	< 0.02	< 0.02	< 0.02

Continued on next page...

Report: Quality Check



OLCC License No. 10087092BDA | ORELAP ID. 4147

545 SW 2nd Street, Corvallis OR. 97333 | 541.257.5002 | services@preelab.com | Preelab.com

For OLCC/OHA Compliance Purposes.

Pesticide Analysis

Quality Control Detail

Pesticide Name	LCS	Expected Recovery (%)	Actual Recovery (%)	Pass Criteria (%)
Abamectin	●	100.00	97.54	60 - 140
Acephate	●	100.00	97.86	60 - 140
Acequinocyl	●	100.00	102.31	60 - 140
Acetamiprid	●	100.00	101.34	60 - 140
Aldicarb	●	100.00	107.48	60 - 140
Azoxystrobin	●	100.00	98.73	60 - 140
Bifenazate	●	100.00	102.36	60 - 140
Bifenthrin	●	100.00	96.84	60 - 140
Boscalid	●	100.00	101.27	60 - 140
Carbaryl	●	100.00	99.59	60 - 140
Carbofuran	●	100.00	100.05	60 - 140
Chlorantraniliprole	●	100.00	100.80	60 - 140
Chlorfenapyr	●	100.00	110.85	60 - 140
Chlorpyrifos	●	100.00	100.85	60 - 140
Clofentezine	●	100.00	98.22	60 - 140
Cyfluthrin	●	100.00	86.17	60 - 140
Cypermethrin	●	100.00	99.47	60 - 140
Daminozide	●	100.00	96.61	60 - 140
Diazinon	●	100.00	99.66	60 - 140
Dichlorvos	●	100.00	99.72	60 - 140
Dimethoate	●	100.00	101.68	60 - 140
Ethoprophos	●	100.00	102.48	60 - 140
Etofenprox	●	100.00	99.98	60 - 140
Etoxazole	●	100.00	101.61	60 - 140
Fenoxycarb	●	100.00	102.49	60 - 140
Fenpyroximate	●	100.00	106.95	60 - 140
Fipronil	●	100.00	102.87	60 - 140
Flonicamid	●	100.00	98.70	60 - 140
Fludioxonil	●	100.00	102.63	60 - 140
Hexythiazox	●	100.00	99.64	60 - 140
Imazalil	●	100.00	95.93	60 - 140
Imidacloprid	●	100.00	99.84	60 - 140
Kresoxim-methyl	●	100.00	101.93	60 - 140

Continued on next page...

Report: Quality Check



OLCC License No. 10087092BDA | ORELAP ID. 4147

545 SW 2nd Street, Corvallis OR. 97333 | 541.257.5002 | services@preelab.com | Preelab.com

For OLCC/OHA Compliance Purposes.

Pesticide Analysis

Quality Control Detail

Pesticide Name	LCS	Expected Recovery (%)	Actual Recovery (%)	Pass Criteria (%)
Malathion	•	100.00	103.25	60 - 140
Metalaxyl	•	100.00	100.51	60 - 140
Methiocarb	•	100.00	98.31	60 - 140
Methomyl	•	100.00	98.34	60 - 140
Methyl-Parathion	•	100.00	88.58	60 - 140
MGK-264 I	•	100.00	106.98	60 - 140
MGK-264 II	•	100.00	101.46	60 - 140
Myclobutanil	•	100.00	100.85	60 - 140
Naled	•	100.00	101.83	60 - 140
Oxamyl	•	100.00	100.55	60 - 140
Paclobutrazol	•	100.00	102.39	60 - 140
Permethrin - trans	•	100.00	109.05	60 - 140
Permethrin - cis	•	100.00	100.33	60 - 140
Phosmet	•	100.00	97.49	60 - 140
Piperonyl butoxide	•	100.00	102.84	60 - 140
Prallethrin	•	100.00	105.63	60 - 140
Propiconazole	•	100.00	100.15	60 - 140
Propoxur	•	100.00	101.65	60 - 140
Pyrethrin - Cinerin	•	100.00	98.57	60 - 140
Pyrethrin - Jasmolin	•	100.00	130.95	60 - 140
Pyrethrin - Pyrethrins	•	100.00	96.10	60 - 140
Pyridaben	•	100.00	108.16	60 - 140
Spinosyn A	•	100.00	97.79	60 - 140
Spinosyn D	•	100.00	95.38	60 - 140
Spiromesifen	•	100.00	111.34	60 - 140
Spirotetramat	•	100.00	100.82	60 - 140
Spiroxamine	•	100.00	98.32	60 - 140
Tebuconazole	•	100.00	100.85	60 - 140
Thiacloprid	•	100.00	100.11	60 - 140
Thiamethoxam	•	100.00	99.37	60 - 140
Trifloxystrobin	•	100.00	103.94	61 - 140

Definitions

- Limit of Quantitation (LOQ) : The minimum level, concentration, or quantity of a target analyte that can be reported with a specific degree of confidence.
- Method Blank (MB) : A quality control sample that is free of the analyte being measured.
- Laboratory Control Sample (LCS) : A quality control sample with a known amount of the analyte used to demonstrate accuracy.
- Field Duplicate : A second sample collected in the field using the same sampling method as the primary sample.
- Action Limit : Analyte levels set by the state of Oregon (OAR 333-007) indicating that follow-up action is necessary.
- ppm : parts per million, equivalent to 1 µg/g and 1 µg/L or 0.001 mg/g and 0.001 mg/L
- COA : Certificate of Analysis.
- Report Flag (A) : Compound tested over 100% or 1000 mg/g. The test result is within the method uncertainty and instrument result is not above the upper limit of quantitation. Value will be adjusted down to 100% or 1000 mg/kg in the reporting process.
- Report Flag (B) : Blank contamination - The analyte was detected above one-half the reporting limit in an associated blank.
- Report Flag (E) : Compound tested above the upper limit of quantitation.
- Report Flag (Q) : One or more quality control criteria (for example, LCS recovery, surrogate spike recovery) failed.

Calculations

- Cannabinoid Potency :
$$\text{Wet WT\%} = (\text{Exported concentration ppm}) \times (\text{Dilution}) \times (\text{Extraction Vol./Wet wt mg}) \times 100$$
$$\text{Total THC\%} = (\% \text{THCA}) \times 0.877 + (\% \text{THC})$$
$$\text{Total CBD\%} = (\% \text{CBDA}) \times 0.877 + (\% \text{CBD})$$
$$\text{Total THC (Dry WT)\%} = \% \text{ total THC(wet)} / [1 - (\% \text{ moisture}/100)]$$
$$\text{Total CBD (Dry WT)\%} = \% \text{ total CBD(wet)} / [1 - (\% \text{ moisture}/100)]$$
- Percentage Recovery :
$$\% \text{ Rec.} = [(\text{Amount measured}) / (\text{Known amount})] \times 100$$

Disclaimers

- Disposal : All marijuana and hemp products received by PREE will be disposed of following the OLCC's rules for Marijuana Waste Management, regardless of product type, unless PREE is given specific disposal instructions for a product based on test results from state regulatory agencies.

EVIO Labs Portland
14775 SW 74th Ave, Tigard, OR 97224
503-954-2562 / OLCC 010-10046111391 / www.EVIOLabs.com

A4994-03

FREE Labs

010-10087092BDA

Sample ID: P210959-03 METRC Batch #:

Matrix: Extract/Concentrate

Date Sampled: 10/27/21 09:00

Date Accepted: 10/27/21

Batch ID:

Batch Size:

Sampling Method/SOP: SOP.T.20.010

Residual Solvents

Analyte	LOQ	Action Level	Result	Units
Butanes	250	5000 ³	< LOQ	ppm
n-Butane	250	5000	< LOQ	ppm
iso-Butane	250	5000	< LOQ	ppm
Hexanes	174	290 ⁴	< LOQ	ppm
n-Hexane	174	290	< LOQ	ppm
2-Methylpentane	174	290	< LOQ	ppm
3-Methylpentane	174	290	< LOQ	ppm
2,2-Dimethylbutane	174	290	< LOQ	ppm
2,3-Dimethylbutane	174	290	< LOQ	ppm
Pentanes	1400	5000 ⁵	< LOQ	ppm
n-Pentane	1400	5000	< LOQ	ppm
iso-Pentane	1400	5000	< LOQ	ppm
Neopentane	250	5000	< LOQ	ppm
Xylenes	1302	2170	< LOQ	ppm
1,2-Dimethylbenzene	1302	2170	< LOQ	ppm
1,3-Dimethylbenzene	1302	2170	< LOQ	ppm
1,4-Dimethylbenzene	1302	2170	< LOQ	ppm
Xylenes MP	1302	2170	< LOQ	ppm
Ethyl benzene	1302	NA	< LOQ	ppm
2-Propanol (IPA)	1400	5000	< LOQ	ppm
Acetone	1400	5000	< LOQ	ppm
Acetonitrile	246	410	< LOQ	ppm
Benzene	1.2	2	< LOQ	ppm
Methanol	1000	3000	< LOQ	ppm
Propane	250	5000	< LOQ	ppm
Toluene	534	890	< LOQ	ppm
Dichloromethane	360	600	< LOQ	ppm
1,4-Dioxane	228	380	< LOQ	ppm
2-Butanol	1400	5000	< LOQ	ppm
2-Ethoxyethanol	96	160	< LOQ	ppm
Cumene	42	70	< LOQ	ppm
Cyclohexane	2278	3880	< LOQ	ppm
Ethyl acetate	1400	5000	< LOQ	ppm
Ethyl ether	1400	5000	< LOQ	ppm
Ethylene glycol	558	620	< LOQ	ppm
Ethylene oxide	30	50	< LOQ	ppm
Heptane	1400	5000	< LOQ	ppm
Isopropyl acetate	1400	5000	< LOQ	ppm
Tetrahydrofuran	432	720	< LOQ	ppm
Ethanol	1400	NA ⁷	< LOQ	ppm
5-hepten-2-one,6-methyl-	NA	TIC	NA	
5-heptenal,2,6-dimethyl-	NA	TIC	NA	
Water	NA	TIC	NA	

Date/Time Extracted: 10/28/21 11:26

Date/Time Analyzed: 10/29/21 11:23

Analysis Method/SOP: SOP.T.40.031

3 - Total butanes are calculated as sum of n-butanenes (CAS# 106-97-8) and iso-butane (CAS# 75-28-5)

4 - Total hexanes are calculated as sum of n-hexane (CAS# 110-54-3), 2-methylpentane (CAS# 107-83-5), 3-methylpentane (CAS# 96-14-0), 2,2-dimethylbutane (CAS# 75-83-2), 2,3-dimethylbutane (CAS# 79-29-8)

5 - Total pentanes are calculated as sum of n-pentane (CAS# 109-66-0), iso-pentane (CAS# 78-78-4), and neo-pentane (CAS# 463-82-1)

6 - Total xylenes are calculated as 1,2-dimethylbenzene (CAS# 95-47-6), 1,3-dimethylbenzene (CAS# 106-42-3), and 1-4-dimethylbenzene (CAS# 106-42-3)

7 - Ethanol is not regulated under OAR-333-007-0410.

TIC - Tentatively Identified Compound not regulated under OAR-333-007-0410

Results above the action level fail Oregon state testing requirements and will be highlighted RED. LOQ=Limit of Quantitation; PPM=Parts per million; ND=Not detected; NT=Not tested; AC=Above calibration range. PASS/FAIL status based on OAR 333-007.



Kawai Medeiros
Laboratory Manager - 10/29/2021

EVIO Labs Portland
14775 SW 74th Ave, Tigard, OR 97224
503-954-2562 / OLCC 010-10046111391 / www.EVIOLabs.com

Quality Control

Batch: P21J078 - SOP.T.40.031 Solvents

Blank(P21J078-BLK1)			Extracted: 10/28/21 11:26		Analyzed: 10/29/21 11:23		
Analyte	Result	LOQ	Recovery Limits	Analyte	Result	LOQ	Recovery Limits
Butanes	< LOQ	250 (ppm)	< LOQ	n-Butane	< LOQ	250 (ppm)	< LOQ
iso-Butane	< LOQ	250 (ppm)	< LOQ	Hexanes	< LOQ	174 (ppm)	< LOQ
n-Hexane	< LOQ	174 (ppm)	< LOQ	2-Methylpentane	< LOQ	174 (ppm)	< LOQ
3-Methylpentane	< LOQ	174 (ppm)	< LOQ	2,2-Dimethylbutane	< LOQ	174 (ppm)	< LOQ
2,3-Dimethylbutane	< LOQ	174 (ppm)	< LOQ	Pentanes	< LOQ	1400 (ppm)	< LOQ
n-Pentane	< LOQ	1400 (ppm)	< LOQ	iso-Pentane	< LOQ	1400 (ppm)	< LOQ
Neopentane	< LOQ	250 (ppm)	< LOQ	Xylenes	< LOQ	1302 (ppm)	< LOQ
1,2-Dimethylbenzene	< LOQ	1302 (ppm)	< LOQ	1,3-Dimethylbenzene	< LOQ	1302 (ppm)	< LOQ
1,4-Dimethylbenzene	< LOQ	1302 (ppm)	< LOQ	Xylenes MP	< LOQ	1302 (ppm)	< LOQ
Ethyl benzene	< LOQ	1302 (ppm)	< LOQ	2-Propanol (IPA)	< LOQ	1400 (ppm)	< LOQ
Acetone	< LOQ	1400 (ppm)	< LOQ	Acetonitrile	< LOQ	246 (ppm)	< LOQ
Benzene	< LOQ	1.2 (ppm)	< LOQ	Methanol	< LOQ	1000 (ppm)	< LOQ
Propane	< LOQ	250 (ppm)	< LOQ	Toluene	< LOQ	534 (ppm)	< LOQ
Dichloromethane	< LOQ	360 (ppm)	< LOQ	1,4-Dioxane	< LOQ	228 (ppm)	< LOQ
2-Butanol	< LOQ	1400 (ppm)	< LOQ	2-Ethoxyethanol	< LOQ	96 (ppm)	< LOQ
Cumene	< LOQ	42 (ppm)	< LOQ	Cyclohexane	< LOQ	2278 (ppm)	< LOQ
Ethyl acetate	< LOQ	1400 (ppm)	< LOQ	Ethyl ether	< LOQ	1400 (ppm)	< LOQ
Ethylene glycol	< LOQ	558 (ppm)	< LOQ	Ethylene oxide	< LOQ	30 (ppm)	< LOQ
Heptane	< LOQ	1400 (ppm)	< LOQ	Isopropyl acetate	< LOQ	1400 (ppm)	< LOQ
Tetrahydrofuran	< LOQ	432 (ppm)	< LOQ	Ethanol	< LOQ	1400 (ppm)	< LOQ

LCS(P21J078-BS1)			Extracted: 10/28/21 11:26		Analyzed: 10/29/21 11:23		
Analyte	% Recovery	LOQ	Recovery Limits	Analyte	% Recovery	LOQ	Recovery Limits
Butanes	67.2	(ppm)	0-200	n-Butane	74.3	(ppm)	50-150
iso-Butane	60.1	(ppm)	50-150	Hexanes	103	(ppm)	0-200
n-Hexane	108	(ppm)	70-130	2-Methylpentane	106	(ppm)	70-130
3-Methylpentane	105	(ppm)	70-130	2,2-Dimethylbutane	105	(ppm)	70-130
2,3-Dimethylbutane	97.8	(ppm)	70-130	Pentanes	117	(ppm)	0-200
n-Pentane	93.9	(ppm)	70-130	iso-Pentane	87.0	(ppm)	70-130
Neopentane	114	(ppm)	50-150	Xylenes	96.6	(ppm)	0-200
1,2-Dimethylbenzene	96.2	(ppm)	70-130	1,3-Dimethylbenzene	97.8	(ppm)	70-130
1,4-Dimethylbenzene	97.8	(ppm)	70-130	Xylenes MP	96.6	(ppm)	0-200
Ethyl benzene	98.1	(ppm)	70-130	2-Propanol (IPA)	101	(ppm)	70-130
Acetone	94.6	(ppm)	70-130	Acetonitrile	106	(ppm)	70-130
Benzene	112	(ppm)	70-130	Methanol	109	(ppm)	70-130
Propane	76.4	(ppm)	50-150	Toluene	103	(ppm)	70-130
Dichloromethane	112	(ppm)	70-130	1,4-Dioxane	107	(ppm)	70-130



Kawai Medeiros
Laboratory Manager - 10/29/2021

EVIO Labs Portland
 14775 SW 74th Ave, Tigard, OR 97224
 503-954-2562 / OLCC 010-10046111391 / www.EVIOLabs.com

Quality Control

Batch: P21J078 - SOP.T.40.031 Solvents (Continued)

LCS(P21J078-BS1)			Extracted: 10/28/21 11:26		Analyzed: 10/29/21 11:23		
Analyte	% Recovery	LOQ	Recovery Limits	Analyte	% Recovery	LOQ	Recovery Limits
2-Butanol	100	(ppm)	70-130	2-Ethoxyethanol	93.9	(ppm)	70-130
Cumene	105	(ppm)	50-150	Cyclohexane	108	(ppm)	70-130
Ethyl acetate	99.6	(ppm)	70-130	Ethyl ether	107	(ppm)	70-130
Ethylene glycol	72.4	(ppm)	70-130	Ethylene oxide	84.4	(ppm)	50-150
Heptane	107	(ppm)	70-130	Isopropyl acetate	106	(ppm)	70-130
Tetrahydrofuran	99.5	(ppm)	70-130				



Kawai Medeiros
 Laboratory Manager - 10/29/2021